

FLX-1005-01-COA

DIN EN ISO 17034:2017



## Certified Reference Material FLX-1005 - Cement

### Certified Values

Parameter	Mass fraction in % <sup>1)</sup>	Uncertainty in % <sup>2)</sup>	Traceable to
Al <sub>2</sub> O <sub>3</sub>	8,57	0,21	NIST 1881B
CaO	50,08	0,49	NIST 1881B
Fe <sub>2</sub> O <sub>3</sub>	3,05	0,11	NIST 1881B
K <sub>2</sub> O	2,34	0,10	IV CGK10 Lot P2-K688009
MgO	1,21	0,07	BAM Reinstoff Nr. 6A
Mn <sub>2</sub> O <sub>3</sub>	0,137	0,021	NIST 1881B
Na <sub>2</sub> O	1,10	0,08	NIST 1881B
P <sub>2</sub> O <sub>5</sub>	0,129	0,020	IV CGP1 Lot U2-P735193
SiO <sub>2</sub>	29,90	0,37	NIST 1881B
SO <sub>3</sub>	2,95	0,16	NIST 1881B
SrO	0,136	0,022	IV CGSR10 Lot R2-SR691242
TiO <sub>2</sub>	0,360	0,036	NIST 1881B
ZnO	0,021	0,008	IV CGZN1 Lot T2-ZN726673

Table1) Certified Values

1) Certified value traceable to SI unit kg/kg based on ignited material.

2) Total expanded uncertainty  $U_{CRM}$  calculated for a confidence interval of 95% ( $k=2$ ).

The sum of all oxides is **100,06 %**. This includes informational values and excludes LOI.

This certificate is valid, within the uncertainty specified, **until 31.09.2034**, provided the CRM is handled in accordance with instructions given in this certificate. The certification is nullified if the CRM is damaged, contaminated, or otherwise modified.

Bedburg-Hau, **10.09.2024**

**Responsible Reference Materials**

Dr. Rainer Schramm

**Quality Management**

Ch. Winkels-Herding

FLX-1005-01-COA

**Description of the CRM**

This reference material is an industrial product and was taken directly from the production stream. The complete batch was sealed into 30 g bottles. This material is normally used as cement for constructions.

**Intended use**

Calibration and control sample for x-ray fluorescence (XRF) analysis.

**Informational Values**

Parameter	Mass Fraction in % <sup>3)</sup>	Uncertainty
Cl	0,072 <sup>4)</sup>	0,006 <sup>5)</sup>
Cr <sub>2</sub> O <sub>3</sub>	0,008 <sup>6)</sup>	0,005 <sup>7)</sup>
LOI	3,67 <sup>8)</sup>	

Table2) Informational Values

3) Only Informational Value.

4) Based on dried material (1h 105 °C). Values taken from PT FXRV-2024-02

5) Total expanded uncertainty  $U_{s^*}$  calculated for a confidence interval of 95% (k=2).

6) Based on ignited material (1h 950°C). Values taken from characterisation.

7) Total expanded uncertainty  $U_{CRM}$  calculated for a confidence interval of 95% (k=2).

8) LOI=loss on ignition 1 hour at 950°C. Based on original material.

**Instructions for the correct use of the CRM**

This material is moisture sensitive. This material has to be ignited for minimum 1 hour at 950°C prior use. The ignition process must result in a constant weight. The ignited material must be stored in a desiccator not longer than 24h, then reignition might be necessary. The minimum sample quantity for analysis should be 0,5 g. For XRF use, ignited samples should be prepared as a fused bead, e.g. in accordance with ISO 29581-2:2010.

**Storage Information**

The material has to be stored in a dry and clean environment.

**Hazardous situation**

For this material an actual MSDS is available.

**Level of homogeneity**

In accordance with ISO 33405:2024 a homogeneity study was performed. A one-way ANOVA was used to calculate the batch inhomogeneity.

**Stability**

In accordance with ISO 33405:2024 a stability study was performed. As a result, the stability of the material was considered as fit for purpose. The uncertainty of long-term stability was calculated

## Total expanded uncertainty

The total expanded uncertainty  $U_{CRM}$  for a confidence interval of 95% ( $k=2$ ) was calculated by taking into account the uncertainty of characterization  $u_{char}$ , of inhomogeneity  $u_{bb}$  and long-term stability  $u_{lts}$ .

$$U_{CRM} = k \times \sqrt{u_{char}^2 + u_{bb}^2 + u_{lts}^2}$$

## Traceability

All of the certified values derived as part of this testing program have traceability to the reference materials stated in table 1.

## Methods used

The analytical work performed to assess this material was carried out by the FLUXANA laboratory, which works according to ISO/IEC 17025:2018.

In accordance with ISO 17034:2017 and ISO 33405:2024, we use the approach stated in ISO 17034:2017 Chapter 7.12.3. d) value transfer from an RM to a closely matched candidate RM performed using a single measurement procedure performed by one laboratory.

An example for this approach is found in ISO 13528:2022 E.5. Using this approach, samples of the test material that is to be the new reference material are tested along with matching and/or synthetic RMs using a suitable method. The assigned values  $X_{CRM}$  and their uncertainties  $U_{CRM}$  are then derived from a calibration against the certified reference values of the compared RMs. Synthetic RMs are made from pure chemicals by weighing.

Measurement method used: XRF with fusion as sample preparation technique.

This certificate is in conformance with ISO 33401:2024.

[feedback@fluxana.de](mailto:feedback@fluxana.de)

